




Sample name: 20250924012
Description: Mobile phase: Buffer A: 0.1%TFA in H2O
Buffer B: 0.1%TFA in ACN
Gradient: 27~37% B in 20min, Flow rate: 1ml/min
Column: Unitary C18, 4.6*150mm, 5um 100A
Instrument: Agilent HPLC 1100
Instrument: 1100
Injection date: 2025/09/24 10:18:49
Last changed: 2025/09/24 11:57:38



Reviewer: Tom Lin