




SS-31

HPLC report

Sample name: SS-31
Description: Mobile phase: Buffer A: 0.1%TFA in H₂O
Buffer B: 0.1%TFA in ACN
Gradient: 15-25% B in 20min, Flow rate: 1ml/min C
column: Unitary C18, 4.6*150mm, 5um 100A
Instrument: Agilent HPLC 1100
Instrument: 1100 Location: 28
Injection date: 2025/09/25 14:52:09 Injection: 1 of 1
Last changed: 2025/09/25 15:37:25 Injection volume: 10.000
Acq. operator: SYSTEM



Signal: DAD1 A, Sig=220, 4 Ref=360, 100

RT[min]	Type	Width[min]	Area	Height	Area%
5.887	BV R	0.1830	30.6298	2.6479	0.0962
6.406	VB E	0.1739	3.0209	0.2800	0.0095
6.945	BV	0.1851	8.3737	0.7125	0.0263
7.187	VV B	0.1621	3.6417	0.3265	0.0114
7.537	VB	0.1518	4.3797	0.4921	0.0138
7.800	BB B	0.1204	1.4601	0.1927	0.0046
8.344	BV R	0.4412	31686.4746	1070.9902	99.5114
9.716	VV E	0.2927	68.1511	3.4767	0.2140
10.901	VB E	0.2717	23.7322	1.2378	0.0745
12.184	BF	0.2797	12.1820	0.6589	0.0383
		Sum	31842.0457		

Inspector: Mike Zhu

Reviewer: Tom Lin